



▶ **Steritest® Biological Control**

Reliably verifies sterility using one million *B. atrophaeus* spores, the spore most resistant to Ethylene Oxide. Sealed Tyvek® / Plastic pouch containing a Dosimeter®, a spore disk, and an ampoule of sterile culture medium. After the sterilization cycle, the dosimeter gives an immediate indication that the appropriate parameters for sterilization have been met. The ampoule is broken and the pouch incubated for 48 hours to confirm sterility. A color change from blue-green to yellow distinctly indicates a positive culture.

▶ **Biological Control Incubator**

Incorporates a thermostatically controlled 37° C incubator into a table top unit with interior dimensions of only 12" x 6" x 12".

Steritest for Anprolene	AN80	11 units/box
Steritest for EOGas	AN1080	11 units/box
Steritest Incubator	AN810	120V/60Hz or 220V/60Hz